



Microbac Laboratories, Inc., New York Division

CERTIFICATE OF ANALYSIS

J611496

Dryden Central School

Project Name: Lead

Robert Medeiros
P.O. Box 88
Dryden, NY 13053

Project / PO Number: N/A
Received: 09/29/2016 15:15
Reported: 11/11/2016 08:17

Analytical Testing Parameters

Client Sample ID: 101 Class. Sink
Lab Sample ID: J611496-01
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:08

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Metals	Result	Limit	PQL	Units	Note	Prepared	Analyzed
Method: EPA 200.8 Rev 5.4							
Lead	<0.00100	0.0150	0.00100	mg/L		11/04/16 0800	11/08/16 1109

Analytical Testing Parameters

Client Sample ID: 102 Class. Sink
Lab Sample ID: J611496-02
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:10

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Metals	Result	Limit	PQL	Units	Note	Prepared	Analyzed
Method: EPA 200.8 Rev 5.4							
Lead	<0.00100	0.0150	0.00100	mg/L		11/04/16 0800	11/08/16 1109

Analytical Testing Parameters

Client Sample ID: 103 Class. Sink
Lab Sample ID: J611496-03
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:11

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Metals	Result	Limit	PQL	Units	Note	Prepared	Analyzed
Method: EPA 200.8 Rev 5.4							
Lead	0.00118	0.0150	0.00100	mg/L		11/04/16 0800	11/08/16 1110



Microbac Laboratories, Inc., New York Division

CERTIFICATE OF ANALYSIS

J611496

Analytical Testing Parameters

Client Sample ID: 104 Class. Sink
Lab Sample ID: J611496-04
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:13

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4, Lead, 0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1111

Analytical Testing Parameters

Client Sample ID: 105 Class. Sink
Lab Sample ID: J611496-05
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:14

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4, Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1111

Analytical Testing Parameters

Client Sample ID: 106 Class. Sink
Lab Sample ID: J611496-06
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:16

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4, Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1113



Microbac Laboratories, Inc., New York Division

CERTIFICATE OF ANALYSIS

J611496

Analytical Testing Parameters

Client Sample ID: 107 Class. Sink
Lab Sample ID: J611496-07
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:17

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4. Row 2: Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1115

Analytical Testing Parameters

Client Sample ID: 108 Class. Sink
Lab Sample ID: J611496-08
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:20

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4. Row 2: Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1116

Analytical Testing Parameters

Client Sample ID: 109 Class. Sink
Lab Sample ID: J611496-09
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:21

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4. Row 2: Lead, 0.00125, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1117



Microbac Laboratories, Inc., New York Division

CERTIFICATE OF ANALYSIS

J611496

Analytical Testing Parameters

Client Sample ID: Primary Fac. East Toilet
Lab Sample ID: J611496-10
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:23

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4. Row 2: Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1117

Analytical Testing Parameters

Client Sample ID: 111 Class. Sink
Lab Sample ID: J611496-11
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:24

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4. Row 2: Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1118

Analytical Testing Parameters

Client Sample ID: 112 Class. Sink
Lab Sample ID: J611496-12
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:26

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4. Row 2: Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1118



Microbac Laboratories, Inc., New York Division

CERTIFICATE OF ANALYSIS

J611496

Analytical Testing Parameters

Client Sample ID: 113 Class. Sink
Lab Sample ID: J611496-13
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:27

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4, Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1119

Analytical Testing Parameters

Client Sample ID: 114 Class. Sink
Lab Sample ID: J611496-14
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:29

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4, Lead, 0.00234, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1120

Analytical Testing Parameters

Client Sample ID: 115 Class. Sink
Lab Sample ID: J611496-15
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:30

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Method: EPA 200.8 Rev 5.4, Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1120



Microbac Laboratories, Inc., New York Division

CERTIFICATE OF ANALYSIS

J611496

Analytical Testing Parameters

Client Sample ID: 116 Class. Sink
Lab Sample ID: J611496-16
Sample Type: Grab

Collected By: RM-Client
Collection Date: 09/29/16
Collection Time: 05:32

Analyses Subcontracted to: Microbac Laboratories, Inc. - Chicagoland

Table with 8 columns: Metals, Result, Limit, PQL, Units, Note, Prepared, Analyzed. Row 1: Lead, <0.00100, 0.0150, 0.00100, mg/L, 11/04/16 0800, 11/08/16 1124

Laboratory

NY: Microbac Laboratories, Inc., New York Division

Definitions

- AL: Action Level
MDL: Minimum Detection Limit
RL: Reporting Limit
G: Result fails applicable NYS drinking water standards

Cooler Receipt Log

Cooler ID: Default Cooler Temp: °C

Cooler Inspection Checklist

Checklist table with 4 columns: Item, Yes, Item, Yes. Items include Custody Seals Intact, COC/Labels Agree, Received on Ice.

Project Requested Certification(s)

Table mapping Microbac Laboratories, Inc. (New York and Chicagoland) to New York State and Pennsylvania Department of Health/Protection.

Report Comments

Samples were received in proper condition and the reported results conform to applicable accreditation standard unless otherwise noted.

Reviewed and Approved By:

Andrew Canale
Project Manager
11/11/2016 08:17

Go Green: Contact Andrew Canale to set up email reporting and invoicing options .

The data and information on this, and other accompanying documents, represents only the sample(s) analyzed. This report is incomplete unless all pages indicated in the footnote are present and an authorized signature is included.

Microbac Laboratories, Inc. CHAIN OF CUSTODY

Samples must be returned on ice
 MNY Workorder #

Client Information		Billing/Invoices:	
Name:	Dryden Elementary School	Dryden Central School District	
Address:	36 Union Street	PO Box 88	
Contact:	Dryden, New York 13053	Dryden, New York 13053	
Phone:	Rob Medeiros	Rob Medeiros	
Project:	607-280-7799	607-280-7799	
Quote ID:		PO#:	
Rush TAT Bus. Days:	< 2-5 5-7 7-10	Date Req.:	
Release to DOH: Yes			
Email Results: Yes	medeir1@dryden.k12.ny.us		
Fax Results: Yes			

Sample Information				Matrix	
Description/Location	Date	Time	Initial	Type	
DES 055	101 Class. Sink	9/29/2016	5:08A	R.M.	
DES 056	102 Class. Sink	9/29/2016	5:10A	R.M.	
DES 057	103 Class. Sink	9/29/2016	5:11A	R.M.	
DES 058	104 Class. Sink	9/29/2016	5:13A	R.M.	
DES 059	105 Class. Sink	9/29/2016	5:14A	R.M.	
DES 060	106 Class. Sink	9/29/2016	5:16A	R.M.	
DES 061	107 Class. Sink	9/29/2016	5:17A	R.M.	
DES 062	108 Class. Sink	9/29/2016	5:20A	R.M.	

Receiving Info (Lab Use Only)	
Ice:	YES NO
Cooler:	YES NO
Sample Temp:	
Cooler Seal:	YES NO
Pickup:	YES NO
Dropoff:	C W
Accepted?	YES NO
Container Material	
Container Size (in MI)	
Preservative	
Comment/Field Data	

Analysis Requested		Number of Containers for Analysis Requested		Comments	
Lead					
P					
250					
HNO3					

Print Name and Company		Signature		Date/Time	
Sampled:	Rob Medeiros/ Dryden Central School District	<i>[Signature]</i>		9/29/16	3:15p
Received:		<i>[Signature]</i>		9/29/16	1515
Received:					
Received:					
Received:					
Received:					



J611496

Microbac Laboratories (MNY) may be unable to perform a portion of the requested testing in which case we will subcontract the analysis to another accredited laboratory. By signing this document you are attesting that you have been informed by MNY of the intent to subcontract and are in agreement with this action.

3821 Buck Drive
 Corland NY 13045
 Phone:(607)753-3403 Fax:(607)753-3415
 NY #10795, EPA #NY00935

Microbac Laboratories, Inc.


CHAIN OF CUSTODY

Samples must be returned on ice
 MNY Workorder #

Client Information		Billing/Invoice:	
Name: Dryden Elementary School	Dryden Central School District	PO Box 88	
Address: 36 Union Street	Dryden, New York 13053	Rob Medeiros	
Contact: Rob Medeiros	607-280-7799		
Phone: 607-280-7799			
Project:			
Quote ID:		PO#:	
Rush TAT Bus. Days: <2 2-5 5-7 7-10	Date Req.:		
Release to DOH: Yes			
Email Results: Yes	rmedeir1@dryden.k12.ny.us		
Fax Results: Yes			

		Sample Information				Matrix Type
		Description/Location	Date	Time	Initial	
DES 063		109 Class. Sink	9/29/2016	5:21A	RM	
DES 064		Primary Fac. East Toilet	9/29/2016	5:23A	RM	
DES 065		111 Class. Sink	9/29/2016	5:24A	RM	
DES 066		112 Class. Sink	9/29/2016	5:26A	RM	
DES 067		113 Class. Sink	9/29/2016	5:27A	RM	
DES 068		114 Class. Sink	9/29/2016	5:29A	RM	
DES 069		115 Class. Sink	9/29/2016	5:30A	RM	
DES 070		116 Class. Sink	9/29/2016	5:32A	RM	

		Analysis Requested			Receiving Info (Lab Use Only)	
		Number of Containers for Analysis Requested	Comments/Field Data	Accepted?		
Lead						YES NO
P						YES NO
250						C W
HNO3						YES NO

Sampled: Rob Medeiros/ Dryden Central School District	Date/Time: 9/29/16 3:50p	Signature: 	Comments:
Received:			
Received:			
Received:			
Received:			
Received:			

*Microbac Laboratories (MNY) may be unable to perform a portion of the requested testing in which case we will subcontract the analysis to another accredited laboratory.
 By signing this document you are attesting that you have been informed by MNY of the intent to subcontract and are in agreement with this action.*